


Traceability overview




Uncertainty analysis example

Quantity [X _i] ID	Name	Value [X _i]	Sensitivity coefficient (Amplification Factor) C _i		Source	Type (A or B)	Distribution	Sensitivity Coeff	dX _i	dX _i /X _i [%]	[%]	Fraction Total Unc.
X1	Vap	10.06475	1	1	Nmi	B	Normal (k=2)	1	3.52E-03	0.035%	0.035%	79%
					1(a)	B	Normal (k=2)	1	1.36E-03	0.013%	0.013%	31%
X2	alpha	-0.0008	7.14E-04	= (DT*DTcv*Vcv/Vp)*alpha	2(a)	B	Normal (k=2)	7.14E-04	-1.20E-04	15.0%	0.01%	24%
X3	beta	9.5E-07	1.75E-04	= (DP*DPcv*Vcv/Vp)*beta	3(a)	B	Normal (k=2)	1.75E-04	1.90E-07	20.0%	0.004%	8%
X4	ΔT	-1.0	7.58E-04	= (alpha + CTE)*DT	4(a)	B	Normal (k=2)	7.58E-04	2.00E-02	-2.00%	-0.002%	3%
					4(b)	A	Normal (k=2)	7.58E-04	1.40E-01	-14.00%	-0.01%	24%
					4(c)	B	Normal (k=2)	7.58E-04	1.40E-01	-14.00%	-0.01%	24%
X5	ΔTcv	0.1	-8.15E-05	= (alpha + CTE)*DTcv*Vcv/Vp	5(a)	B	Normal (k=2)	-8.15E-05	2.00E-02	20.0%	-0.002%	4%
					5(b)	A	Normal (k=2)	-8.15E-05	1.40E-01	140.00%	-0.01%	26%
					5(c)	B	Normal (k=2)	-8.15E-05	0.00E+00	0.00%	0.000%	0%
X6	ΔP	-200.0	1.90E-04	= beta*DP	6(a)	B	Normal (k=2)	1.90E-04	40.00	-20.00%	-0.004%	9%
					6(b)	B	Normal (k=2)	1.90E-04	20.00	-10.00%	-0.002%	4%
					6(c)	A	Normal (k=2)	1.90E-04	20.00	-10.00%	-0.002%	4%
X7	ΔPcv	14.5	-1.48E-05	= beta*DPcv*Vcv/Vp	7(a)	B	Normal (k=2)	-1.48E-05	40.00	275.86%	-0.004%	9%
					7(b)	B	Normal (k=2)	-1.48E-05	20.00	137.93%	-0.002%	5%
					7(c)	A	Normal (k=2)	-1.48E-05	20.00	137.93%	-0.002%	5%
											0.044%	

	Values	Imperial Units	SI Units	Comments
X1	Vp	63.3010 barrels	10.06475 cubic meters	Prover Volume
X2	alpha	-4.45E-04 vol/vol/degF	-0.0008 vol/vol/degC	Fluid CTE - Used Average Value for the 3 liquids evaluated by Nmi
	density	38.66158 API Degrees	830.74 kg/m ³	Used Average Value for the 3 liquids evaluated by Nmi
X3	beta	6.55E-06 1/psi	9.5E-07 1/kPa	Estimate from API Tables
	Tp	95.0 degF	35.0 degC	Maximum temperature as worst case conditions
	Tm	96.8 degF	36.0 degC	
	Pp	21.0 psi	144.8 kPa	
	Pm	50.0 psi	344.8 kPa	
X8	CTE prover	2.34E-05 vol/vol/degF	4.21E-05 vol/vol/degC	Nil
X9	CTE meter	3.33E-05 vol/vol/degF	5.99E-05 vol/vol/degC	Nil
	CTE	2.34E-05 vol/vol/degF	4.21E-05 vol/vol/degC	
	N	10000 Counts	10000 Counts	Kfactor forced to produce 10,000 Counts.
	Vcv	68.0749 barrels	10.824 cubic meters	See inventory calculation of connecting volume
X5	ΔTcv = Tcvf - Tcvi	0.2 degF	0.1 degC	Change in temperature in connecting volume during test is constrained to be less than 0.1 deg C
X7	ΔPcv = Pcvf - Pcvi	1.0 psi	14.5 kPa	Change in pressure in connecting volume during test is constrained to be less than 14.5 kPa
Known	Tbase	60 degF	15.6 degC	
Known	Pe	14.5 psi	100.0 kPa	
Known	K	0.00633 bbl/count	0.001006 cm ³ /count	
X4	ΔT = (Tp - Tm)	-1.8 degF	-1.0 degC	Difference in temperature between the meter and the prover is constrained to be less than 1 deg C
X6	ΔP = (Pp - Pm)	-29.0 psi	-200.0 kPa	Difference in pressure between the meter and the prover is constrained to be less than 200 kPa
	Change Vcv	-0.0056 barrels	-0.0010 cubic meters	
	Vm	6.34E-01		


Major contributors to uncertainty

- Prover base volume
- Coefficient of thermal expansion for the liquid
- Delta temperature between prover and meter under test
- Delta temperature for connecting volume




NVLAP Certified Uncertainties

- 10 to 750 m³/hr
 - Small volume prover 0.03%
 - Turbine master meter 0.04%
- 150 to 2200 m³/hr
 - Ball prover 10 m³ 0.04%
 - Ball prover 3.3 m³ 0.07%
 - One master meter 0.09%
- 600 to 3900 m³/hr
 - Two master meters 0.08%



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Cameron Measurement Systems
Caldon Ultrasonics Technology Center
1000 McClaren Woods Drive
Coraopolis, PA 15108-7766
Mr. Bobbie Griffith
Phone: 724-273-9134 Fax: 724-273-9301
E-mail: bobbie.griffith@c-a-m.com

CALIBRATION LABORATORIES

NVLAP LAB CODE 200813-0
Scope Revised: 2011-08-19

MECHANICAL

NVLAP Code: 20/M05
Flow Rate (Hydrocarbon Fluids Only)^{1,2}

Range in m ³ /h	Best Uncertainty (±) in % ^{1,2}	Remarks
10 to 750	0.03	Brooks Small Volume Prover
10 to 750	0.04	One Master Meter
150 to 2200	0.04	10 Cubic Meter Prover Volume
50 to 200	0.07	3.3 Cubic Meter Prover Volume
300 to 2000	0.09	One Master Meter
600 to 3900	0.08	Two Master Meters

1. Represents an expanded uncertainty using a coverage factor, *k* = 2, at an approximate level of confidence of 95%.

2. The laboratory performs calibrations of pulse generating flow meters.


3. The laboratory performs volumetric flow calibrations only (not mass flow).

2011-07-01 through 2012-06-30
Effective dates

Sally S. Bruce
For the National Institute of Standards and Technology

Page 1 of 1
NVLAP-218 (REV. 2004-10-31)

Comparison of ISO17025 capabilities




Key standards and documentation

- International Standards Organisation ISO17025:2005
General requirements for the competence of testing and calibration laboratories
- Joint Committee for Guides in Metrology JCGM 100:2008
(GUM 1995 with minor corrections) Evaluation of measurement data — Guide to the expression of uncertainty in measurement
- API Manual of Petroleum Measurement Standards
 - Especially the subchapters of Chapter 4 - Proving
- Selected ISO and ASTM standards

Calibration laboratory management system

- Independent quality system accredited for conformance with ISO17025
- Calibration laboratory quality manual comprising 28 documents
- Plus 26 documented operational procedures



Document Number	Document Title	Rev. No.	Effective Date	Prepared By	Pages	Review Period	Reviewer	Review Date	Review Signature
Annex A	Use and Control of the NVLAP Term and Logo	0	02/19/10	Coates	2	Two Years	QA Manager		
CLM 4.1	Flow Laboratory Organization	2	05/11/10	Outfth	3	Two Years	QA Manager		
CLM 4.2	Flow Laboratory Management System	2	07/11/11	Outfth	2	Two Years	QA Manager		
CLM 4.3	Flow Laboratory Document Control	0	11/9/09	Coates	4	Two Years	QA Manager		
CLM 4.4	Flow Laboratory Contract Review	1	03/20/10	McCarter	2	Two Years	QA Manager		
CLM 4.5	Subcontractor Control	1	12/12/07	McCarter	1	Two Years	QA Manager		
CLM 4.6	Purchasing Services and Supplies	2	01/20/10	McCarter	2	Two Years	QA Manager		
CLM 4.7	Customer Service	1	12/13/07	McCarter	1	Two Years	QA Manager		
CLM 4.8	Customer Complaints	2	02/20/10	Coates	1	Two Years	QA Manager		
CLM 4.9	Nonconformance Control	2	01/28/10	Coates	2	One Year	QA Manager		
CLM 4.10	Continual Improvement	0	08/07/07	McCarter	1	One Year	QA Manager		
CLM 4.11	Corrective Action	0	08/20/10	McCarter	2	One Year	QA Manager		
CLM 4.12	Preventative Action	0	1/21/10	McCarter	0	One Year	QA Manager		
CLM 4.13	Control of Records	0	08/08/07	McCarter	2	Two Years	QA Manager		
CLM 4.14	Flow Lab Internal Audit Procedures	0	12/13/07	McCarter	3	Two Years	QA Manager		
CLM 4.15	Flow Lab Management Review Procedures	0	12/13/07	McCarter	1	Two Years	QA Manager		
CLM 5.1	General	0	01/15/04	Coates	1	Two Years	Flow Director		


Master meter long term stability

- Master meters are calibrated before use
- Long term stability gives a good indication that everything is working as expected




Master meter long term stability

Time Vs MF bar For 10 inch 280C MM1 for the following Re



Temperature control




Intercomparison with NEL (UK)

- ISO 17025 states that “Participation in a suitable programme of interlaboratory comparisons is required where possible”
- Flow intercomparisons involving multiple laboratories are rare and tend to involve national measurement institutes
- However, NEL in the UK had been designated as ‘pilot lab’ in the recently completed Liquid Hydrocarbon Flow Key Comparison CCM-FF-K2, carried out under the auspices of the International Committee of Weights and Measures (CIPM)

Intercomparison with NEL (UK)


- Other participants in CCM-FF-K2 were:
 - CMS (Chinese Taipei), NMIJ (Japan), NMI VSL (Netherlands), FORCE (Denmark), and SP (Sweden)
- Therefore by carrying out a bi-lateral intercomparison with NEL, a successful result would demonstrate metrological equivalence with a broad group of national metrology institutes

Key Comparison CCM-FF-K2 results



Intercomparison package


- An intercomparison package was assembled using two 6-inch Caldon 8-path meters
- A flow conditioning upstream section was used to ensure absolutely minimal installation effects



- | | |
|---------|---|
| A and G | – 5D straight length (matched schedule with internal welds ground flush) |
| B | – Perforated plate flow conditioner (CPA) |
| C | – 15D straight length (matched schedule with internal welds ground flush) |
| D & F | – 280Ci ultrasonic flow meters |
| E | – 3D 150# to 300# flange adaptor spool |

Intercomparison package

Cameron




NEL, UK




Intercomparison package

- Primary comparison was carried out using kerosene substitute (Exxsol D80) over a flow range of 100 to 600 m³/hr in both facilities
 - Cameron lab tests vs ball prover
 - NEL tests versus turbine secondary standards
- A secondary comparison was also carried out using the NEL water flow facility gravimetric standard
- Comparisons were made at overlapping Reynolds numbers

Intercomparison results on kerosene



Intercomparison results vs water



Intercomparison results

- The results from both meters and both NEL facilities (oil and water) demonstrated metrological equivalence with the Cameron laboratory
- The closest and most linear agreement was actually found in the case of the water comparison, suggesting that the difference in the oil calibration were in part due to the curve fitting to the NEL secondary standard turbines

SUMMARY

- The Cameron laboratory as constructed exceeds the design expectations
- It offers a capability that surpasses that of the third-party facilities previously used
- The uncertainty has been validated by VSL, NVLAP and the NEL intercomparison
- Use of well understood and reliable system components and methods have resulted in a high availability – the laboratory is in almost constant use

Thank you



gregor.brown@c-a-m.com